

TEST REPORT

(This Report is issued subject to the terms & conditions set out below)

Setsco Services Pte Ltd 18 Teban Gardens Crescent Singapore 608925 Tel : (65) 6566 7777 Fax: (65) 6566 7718 www.setsco.com Business Reg. No. 196900269D

Your Ref: -Our Ref: FB-8500159479/SS/4 Date: 18/05/2021 Page 1 of 3 Subject : Differential pressure test for Granzilla Aircare 3 ply disposable face mask (Black/White) submitted by Grantech Pte Ltd on 26/04/2021 Purpose The purpose of the test is to measure the differential pressure of a medical 2 face mask **Tested For** : GRANTECH PTE LTD 629 Aljunied Road #04-17 Cititech Industrial Building Singapore 389838 Attn: Mr SY Ong Method of Test : BS EN 14683 : 2019 Medical Face Masks - Requirements and Test Methods Clause 5.2.3 Annex C Method for determination of breathability (differential pressure) **Description of** : 50 pieces of Granzilla Aircare 3 ply disposable face mask (Black/White) were Samples received (Refer to Photograph 1) Lotcode 20210104 Model SML-A21 Meltblown Used HSL200826

Khairuddin **Testing Officer**

Salim Suwignjo Assistant Manager (CMTD) Construction Technology Division

- (2) SETSCO agrees to use reasonable diligence in the performance of the Services but no warranties are given and done may be implied directly or indirectly relating to the Services, the Report or the facilities of SETSCO.
- (3) SETSCO shall under no circumstances be liable to the Client or its agents, servants or representatives, in contract, tort (including negligence or breach of statutory duty) or otherwise for
- any direct or indirect loss or damage suffered by the Client, its agents, servants or representative howsoever arising or whether connected with the Services provided by SETSCO herein. **ac-MRA**
- (4) The Report may not be used in any publicity material without the written consent of SETSCO.
- (5) The Report shall not be reproduced in part or in full unless approval in writing has been given by SETSCO.



Accredited Laboratory SAC-SINGLAS LA-1993-0051-C LA-1993-0067-G LA-1994-0068-A

LA-1987-0001-B

LA-1998-0144-D LA-2000-0181-F LA-2012-0519-E LA-1987-0001-B-1

⁽¹⁾ The Report is prepared for the sole use of the Client and is prepared based upon the Item submitted, the services required by the Client and the conditions under which the Services are performed by SETSCO. The Report is not intended to be representative of similar or equivalent Services on similar or equivalent items. The Report does not constitute an endorsement by SETSCO of the item.



Page 2 of 3

FB-8500159479/SS/4

Test Result : Differential Pressure

Sample Reference	Granzilla Aircare 3 ply disposable face mask (Black/White)					
Conditioning		17/05/2021 From 1000hrs to 1400hrs @ 21°C, 85%RH				
Date of Test		17/05/2021				
Test Condition		23°C, 52%RH				
Air Flow (L/min)		8				
Tested Area (cm ²)		4.9				
Type of Mask		Multi Material Mask				
Specimen Reference		S1	S2	S3	S4	S5
Differential Pressure, ΔP (Pa/cm²)	Top 1	40.61	43.27	40.82	43.27	40.00
	Top 2	44.08	45.71	41.02	41.84	43.67
	Тор 3	46.93	42.45	47.35	41.63	44.08
	Bottom 1	42.04	44.08	38.78	42.45	40.82
	Bottom 2	41.22	46.73	40.00	40.00	45.31
	Bottom 3	45.10	47.96	48.16	45.71	45.31
	Average	43.33	45.03	42.69	42.48	43.20
Requirement BS EN 14683 Clause 5.2.7 Summary of Performance Requirements		Differential Pressure, (Pa/cm ²)				
		Туре I < 40				
		Type II < 40				
		Type IIR < 60				

h

R



FB-8500159479/SS/4

1

1

Page 3 of 3



Photograph 1 Granzilla Aircare 3 ply disposable face mask (Black/White) as received



Photograph 2 Granzilla Aircare 3 ply disposable face mask (Black/White) Outer (Top) and Inner (Bottom)