

TEST REPORT

(This Report is issued subject to the terms & conditions set out below)

SetSCO Services Pte Ltd
 18 Teban Gardens Crescent
 Singapore 608925
 Tel : (65) 6566 7777
 Fax: (65) 6566 7718
 www.setsco.com
 Business Reg. No. 196900269D

Your reference No. : -
 Our reference No. : FB8500163443

Page 1 of 3

SYNTHETIC BLOOD PENETRATION TEST

Date: 07/06/2021

Tested For: **Grantech Pte Ltd**
 629 Aljunied Road
 Unit 04-17
 Cititech Industrial Building
 Singapore 389838
Attention : SY Ong

Test Method: ASTM F1862 / ISO22609 Resistance of Medical Face Masks to Penetration by Synthetic Blood (Horizontal Projection of Fixed Volume at a Known Velocity).

Date Sample Received: 31/05/2021

Date Sample Analyzed: 04/06/2021

Sample Description: Granzilla Aircare 3 Ply Disposable Face Mask (Black/White) (LOT : LOT21042021)


Results:

Lab Sample ID No.	Pressure	Number of Test Articles Tested	Number of Test Articles Passed ¹
FB8500163443	160 mmHg (21.3 kPa)	32	32

Remarks:

The tested results apply only to the samples as received by the laboratory.

¹As per ASTM F1862 Level III face masks, an acceptable quality limit of 4.0% is met for a normal single sampling plan when ≥ 29 out of 32 test articles show passing results.


 DR WU JIEN
 HOD MICROBIOLOGY

Terms & Conditions:

- (1) The Report is prepared for the sole use of the Client and is prepared based upon the Item submitted, the services required by the Client and the conditions under which the Services are performed by SETSCO. The Report is not intended to be representative of similar or equivalent Services on similar or equivalent items. The Report does not constitute an endorsement by SETSCO of the item.
- (2) SETSCO agrees to use reasonable diligence in the performance of the Services but no warranties are given and none may be implied directly or indirectly relating to the Services, the Report or the facilities of SETSCO.
- (3) SETSCO shall under no circumstances be liable to the Client or its agents, servants or representatives, in contract, tort (including negligence or breach of statutory duty) or otherwise for any direct or indirect loss or damage suffered by the Client, its agents, servants or representative howsoever arising or whether connected with the Services provided by SETSCO herein.
- (4) The Report may not be used in any publicity material without the written consent of SETSCO.
- (5) The Report shall not be reproduced in part or in full unless approval in writing has been given by SETSCO.

BIOLOGICAL & CHEMICAL TECHNOLOGY DIVISION

Your reference No. : -

Our reference No. : FB8500163443

Page 2 of 3

SYNTHETIC BLOOD PENETRATION TEST

Test Items Granzilla Aircare 3 Ply Disposable Face Mask (Black/White) (LOT : LOT21042021)	Test Results (Blood penetration @ 160 mmHg)
Sample #1	None Observed
Sample #2	None Observed
Sample #3	None Observed
Sample #4	None Observed
Sample #5	None Observed
Sample #6	None Observed
Sample #7	None Observed
Sample #8	None Observed
Sample #9	None Observed
Sample #10	None Observed
Sample #11	None Observed
Sample #12	None Observed
Sample #13	None Observed
Sample #14	None Observed
Sample #15	None Observed
Sample #16	None Observed

Remarks:

The tested results apply only to the samples as received by the laboratory.

¹As per ASTM F1862 Level III face masks, an acceptable quality limit of 4.0% is met for a normal single sampling plan when ≥29 out of 32 test articles show passing results.


DR WU JIEN
HOD MICROBIOLOGY

BIOLOGICAL & CHEMICAL TECHNOLOGY DIVISION

Your reference No. : —
 Our reference No. : FB8500163443

SYNTHETIC BLOOD PENETRATION TEST

Test Items Granzilla Aircare 3 Ply Disposable Face Mask (Black/White) (LOT : LOT21042021)	Test Results (Blood penetration @ 160 mmHg)
Sample #17	None Observed
Sample #18	None Observed
Sample #19	None Observed
Sample #20	None Observed
Sample #21	None Observed
Sample #22	None Observed
Sample #23	None Observed
Sample #24	None Observed
Sample #25	None Observed
Sample #26	None Observed
Sample #27	None Observed
Sample #28	None Observed
Sample #29	None Observed
Sample #30	None Observed
Sample #31	None Observed
Sample #32	None Observed

Remarks:

The tested results apply only to the samples as received by the laboratory.

¹As per ASTM F1862 Level III face masks, an acceptable quality limit of 4.0% is met for a normal single sampling plan when ≥ 29 out of 32 test articles show passing results.



DR WU JIEN
 HOD MICROBIOLOGY

BIOLOGICAL & CHEMICAL TECHNOLOGY DIVISION